Resmed

AirCurve™ 11 ASV

AirCurve 11 ASV Prescription Form

Indication for use:

The AirCurve 11 ASV* device is indicated for the treatment of Obstructive Sleep Apnoea (OSA) in patients weighing more than 66 lbs (30 kg). ASV and ASVAuto modes are also indicated for the treatment of central and/or mixed apnoeas, or periodic breathing. The AirCurve 11 ASV device is intended for use in the hospital and home.

Patient details

Name

Date of birth (DD-Mon-YYYY)

Medicare card number

Phone

Individual reference number

Address

Therapy mode

Therapy Mode ◯ **ASV** ◯ **ASVAuto** ◯ **CPAP**

ASV and ASVAuto therapy settings

Parameter	Prescribed value	Unit	Device default	Device range	OFFICE USE Check 1	OFFICE USE Check 2
Fixed EPAP (ASV mode)						
EPAP:		cmH_2O	4	4 to 15, in increments of 0.2		
EPAP Range (ASVAuto mode)						
Min. EPAP:		cmH_2O	4	4 to 15, in increments of 0.2		
Max. EPAP:		cmH_2O	15	"Min. EPAP" to 15, in increments of 0.2		
Pressure support range (ASV and ASVAuto mode)						
Min. PS:		cmH_2O	3	0 to 6, in increments of 0.2		
Max. PS:		cmH_2O	15	5 to 20, in increments of 0.2		

CPAP therapy settings

Parameter	Prescribed value	Unit	Device default	Device range	OFFICE USE Check 1	OFFICE USE Check 2
Set pressure:		cmH_2O	10	4 to 20, in increments of 0.2	☐	☐

Supplemental oxygen settings

Does the patient need supplemental oxygen with their device? ◯ **Yes** ◯ **No**

Parameter	Prescribed value	Unit	Device default	Device range	OFFICE USE Check 1	OFFICE USE Check 2
Flow rate:		L/min	-	Available only when supplemental oxygen selected is "Yes". Up to 15L/min in all modes	☐	☐

Continue to next page

*ASV therapy is contraindicated in patients with chronic, symptomatic heart failure (NYHA 2-4) with reduced left ventricular ejection fraction (LVEF ≤ 45%) and moderate to severe predominant central sleep apnoea.

Resmed

AirCurve™ 11 ASV

AirCurve 11 ASV Prescription Form

Comfort settings

Ramp: ◯ On ◯ Off

Parameter	Prescribed value	Unit	Device default	Device range	OFFICE USE Check 1	OFFICE USE Check 2
Ramp time:		min	20	Available only when ramp time selected is "On".5 to 45, in increments of 5	☐	☐
Start EPAP:		cmH_2O	4	Available in ASV and ASVAuto mode only. 4 to "EPAP" or "Min. EPAP", in increments of 0.2		
Start pressure:		cmH_2O	4	Available in CPAP mode only. 4 to "Set pressure", in increments of 0.2	☐	☐

Mask type:

◯ **Full face** ◯ **Nasal** ◯ **Pillow** ◯ **Other**

If other, please specify

Prescriber Details

Name

Phone

Address

Email

Provider number

Date (DD-Mon-YYYY)

Signature

Reset form | Print

MANUFACTURER
Resmed Pty Ltd
1 Elizabeth Macarthur Drive
Bella Vista NSW 2153 Australia