



Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorized Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Product: Standard and SlimLine Tubing Wrap

Intended Use:

The Standard and SlimLine Tubing Wrap are accessories intended to reduce condensation and temperature loss in ResMed air tubing when connected to a device. It is not a single use device.

Classification: I according to Rule 2

EMDN: Z1203019080 Various Instruments for Anaesthesia and Pulmonary Ventilation Support - Hardware Accessories

Conformity Assessment Route: Annex II and III, Regulation EU 2017/745

Basic UDI-DI: 619498EC0786V

Common Specification: N/A

We herewith declare that the above mentioned products are in conformity with the Council Regulation 2017/745 for medical devices.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer. This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

SRN: AU-MF-000011753

Signed at Sydney, Australia on: 03 October 2023

DocuSigned by:

Nicole Wilson

F9F55744DEA64A8...

Nicole Wilson
Person Responsible for Regulatory Compliance (PRRC)
ResMed Pty. Ltd.

EC078.1

First issued: 03 October 2023