



Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorised Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Product: Gecko Nasal Pad

Intended Use: The Gecko Nasal Pad is intended to provide a mechanical barrier between the patient's mask and face. It is suitable for use with full face and nasal masks.

Classification: I according to Rule 4

GMDN: 57814 CPAP/BPAP face mask, reusable

Conformity Assessment Route: Annex VII, 93/42/EEC

We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

Signed at Sydney, Australia on: 11 December 2019

A handwritten signature in cursive script, appearing to read "Johanna Wright".

Johanna Wright
Director of Regulatory Affairs
ResMed Pty. Ltd.

EC126

First issued: 23 June 2011