



TÜV SÜD Product Service

GmbH

Ridlerstraße 65

80339 München

Declaration of Conformity

Manufacturer: **Authorised Representative: Notified Body:**

ResMed Pty. Ltd. ResMed SAS

1 Elizabeth Macarthur Drive Parc Technologique de Lyon Bella Vista 292 Allée Jacques Monod NSW 2153

Australia France Germany

Product: Air10 Oximeter Adapter

Intended Use: The Air10 Oximeter Adapter is intended to be used with AirSense} 10 / AirCurve} 10

69791 Saint Priest Cedex

devices. The oximeter adapter connects a NONIN Xpod) oximeter to a device. It is

intended for use in the home and hospital/instituional environment.

Classification: IIa according to Rule 9

GMDN: 60711 Home CPAP unit

Conformity Assessment Route: Annex II (excluding Section 4), 93/42/EEC

We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

EC Certificate Number: G1 049861 0158

Signed at Sydney, Australia on: 20 August 2020

Johanna Wright

Director of Regulatory Affairs

ResMed Pty. Ltd.