

Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorised Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Notified Body:

TÜV SÜD Product Service
GmbH
Ridlerstraße 65
80339 München
Germany

Product: ClimateLineAir Oxy

Intended Use: The ClimateLineAir Oxy heated air tubing is designed for use with AirSense™ 10/AirCurve™ 10 and Lumis™ devices. It is intended for single patient use in the home environment and re-use in a hospital/institutional environment.

Classification: IIa according to Rule 9

GMDN: 37705 Breathing circuit, ventilator, reusable

Conformity Assessment Route: Annex II (excluding Section 4), 93/42/EEC


We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

EC Certificate Number: G1 049861 0158

Signed at Sydney, Australia on: 11 December 2019



Johanna Wright
Director of Regulatory Affairs
ResMed Pty. Ltd.

EC182b

First issued: 13 February 2015