



Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorised Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Product:

Astral DC-DC Adapter

Intended Use:

The Astral DC-DC Adapter provides an additional method of operating an Astral 100 or Astral 150 Ventilation device where AC mains power is not available. The Adapter plugs directly into a standard DC socket (eg, an automotive cigarette lighter socket) to power the Astral device and charge the internal battery. The Adapter can be used with power sources ranging from 12V to 24V DC.

Classification:

I according to Rule 12

GMDN:

47083 Portable ventilator, electric

Conformity Assessment Route: Annex VII, 93/42/EEC

We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

Signed at Sydney, Australia on: 11 December 2019

A handwritten signature in cursive script, appearing to read "Johanna Wright".

Johanna Wright
Director of Regulatory Affairs
ResMed Pty. Ltd.

EC129a

First issued: 20 February 2019