

Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorised Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Notified Body:

TÜV SÜD Product Service
GmbH
Ridlerstraße 65
80339 München
Germany

Product:

ResMed Connectivity Module (RCM1)

Intended Use:

RCM is intended to be used in the home environment, for the collection and transmission of respiratory data to AirView. RCM will not control any clinical devices, nor provide interpretation of data.

RCM is not intended for use on an aircraft.

Classification:

Ila according to Rule 9

GMDN:

36862 Patient monitoring system module, interfacing

Conformity Assessment Route: Annex II (excluding Section 4), 93/42/EEC


We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC, Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment and Directive 2014/53/EU.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

EC Certificate Number: G1 049861 0158

Signed at Sydney, Australia on: 11 December 2019



Johanna Wright
Director of Regulatory Affairs
ResMed Pty. Ltd.

EC169a

First issued: 29 April 2016