



Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorized Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Product: Astral DC-DC Adapter

Intended Use:

The Astral DC-DC Adapter provides an additional method of operating an Astral 100 or Astral 150 Ventilation device where AC mains power is not available. The Adapter plugs directly into a standard DC socket (eg, an automotive cigarette lighter socket) to power the Astral device and charge the internal battery. The Adapter can be used with power sources ranging from 12V to 24V DC.

Classification: I according to Rule 13

EMDN: Z1203019080 Various Instruments for Anaesthesia and Pulmonary Ventilation Support - Hardware Accessories

Conformity Assessment Route: Annex II and III, Regulation EU 2017/745

Basic UDI-DI: 619498EC1806M

Common Specification: N/A

We herewith declare that the above mentioned products are in conformity with the Council Regulation 2017/745 for medical devices.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer. This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

SRN: AU-MF-000011753

Signed at Sydney, Australia on: 29 June 2023

DocuSigned by:

Nicole Wilson

F9F35744DEA04A0...

Nicole Wilson
Person Responsible for Regulatory Compliance (PRRC)
ResMed Pty. Ltd.

EC180d.1

First issued: 15 May 2023