

Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorised Representative:

ResMed SAS
Pac Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Notified Body:

TÜV SÜD Product Service
GmbH
Ridlerstraße 65
80339 München
Germany

Product: AirFit F20 NV

Intended Use: The AirFit F20 Non-Vented full face mask is a patient interface to deliver non-invasive positive pressure ventilation. It is intended to be used as an accessory to ventilators that have adequate alarms and safety systems for ventilator failure, to administer continuous or intermittent ventilatory support.

The AirFit F20 Non-Vented full face mask is:

- to be used by patients weighing more than 30 kg
- intended for single-patient re-use in the home environment and/or multi-patient re-use in the hospital/institutional environment.

Classification: IIa according to Rule 2

GMDN: 57814 CPAP/BPAP face mask, reusable

Conformity Assessment Route: Annex II (excluding Section 4), 93/42/EEC


We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

EC Certificate Number: G1 049861 0158

Signed at Sydney, Australia on: 21 May 2021



Johanna Wright
Director of Regulatory Affairs
ResMed Pty. Ltd.

EC201

First issued: 21 May 2021