



0123

Declaration of Conformity

| Manufacturer: | Authorized Representative: | Notified Body: |
|---|---|--|
| ResMed Pty. Ltd. 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia | ResMed SAS Parc Technologique de Lyon 292 Allée Jacques Monod 69791 Saint Priest Cedex France | TÜV SÜD Product Service GmbH Ridlerstraße 65 80339 München Germany |

Product: Quattro Air NV

Intended Use:

The Quattro Air NV is a non-invasive accessory used for channelling air-flow (with or without supplemental oxygen), intended to be used with active-exhaust-valve ventilators, to provide ventilatory assistance to patients with respiratory insufficiency and respiratory failure.

The Quattro Air NV is:

- to be used by patients weighing (>66 lb/30 kg) requiring non-life support ventilatory assistance
- intended for single patient re-use in the home environment and/or multi-patient re-use in the hospital/institutional environment

Classification: IIa according to Rule 2

EMDN: R0301010202 NIV Masks

Conformity Assessment Route: Annex IX (excluding Chapter II), Regulation EU 2017/745

Basic UDI-DI: 619498EC1476P

Common Specification: N/A

We herewith declare that the above mentioned products are in conformity with the Council Regulation 2017/745 for medical devices.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer. This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

EC Certificate Number: G15 049861 0261 Rev. 01

SRN: AU-MF-000011753

Signed at Sydney, Australia on: 18 February 2026

DocuSigned by:
Nicole Wilson
F9F55744DEA64A0...

Nicole Wilson
Person Responsible for Regulatory Compliance (PRRC)
ResMed Pty. Ltd.

EC147.1

First issued: 12 August 2025