



0123

## Declaration of Conformity

---

<b>Manufacturer:</b>	<b>Authorized Representative:</b>	<b>Notified Body:</b>
ResMed Pty. Ltd. 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia	ResMed SAS Parc Technologique de Lyon 292 Allée Jacques Monod 69791 Saint Priest Cedex France	TÜV SÜD Product Service GmbH Ridlerstraße 65 80339 München Germany

---

**Product:** AirFit N20 and AirFit N20 For Her

**Intended Use:**

The Air Fit N20 / Air Fit N20 For Her Nasal Mask System is a non-invasive accessory used for channelling air-flow to a patient from a positive airway pressure device such as Continuous Positive Airway Pressure (CPAP) or bilevel system.

The Air Fit N20 / Air Fit N20 For Her Nasal Mask System is:

- to be used by patients (weighing >66 lb/30 kg) for whom positive airway pressure therapy has been prescribed.
- intended for single patient re-use in the home environment and multi-patient re-use in the hospital/institutional environment.

**Classification:** IIa according to Rule 2

**EMDN:** R0301010201 CPAP Masks

**Conformity Assessment Route:** Annex IX (excluding Chapter II), Regulation EU 2017/745

**Basic UDI-DI:** 619498EC1736Q

**Common Specification:** N/A

---

We herewith declare that the above mentioned products are in conformity with the Council Regulation 2017/745 for medical devices.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer. This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

**EC Certificate Number:** G15 049861 0261 Rev. 01

**SRN:** AU-MF-000011753

Signed at Sydney, Australia on: 01 May 2026

DocuSigned by:  
*Nicole Wilson*  
F9F55744DEA64A0...

---

Nicole Wilson  
Person Responsible for Regulatory Compliance (PRRC)  
ResMed Pty. Ltd.

**EC173.1**

First issued: 01 May 2026