



0123

## Declaration of Conformity

---

<b>Manufacturer:</b>	<b>Authorized Representative:</b>	<b>Notified Body:</b>
ResMed Pty. Ltd. 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia	ResMed SAS Parc Technologique de Lyon 292 Allée Jacques Monod 69791 Saint Priest Cedex France	TÜV SÜD Product Service GmbH Ridlerstraße 65 80339 München Germany

---

**Product:** HumidAir

**Intended Use:**

The humidifier is intended for single patient use in the home environment and re-use in a hospital/institutional environment.

It is intended for use with the AirSense/AirCurve Series 10 CPAP or Bilevel devices.

**Classification:** IIa according to Rule 9

**EMDN:** Z1203019080 Various Instruments for Anaesthesia and Pulmonary Ventilation Support - Hardware Accessories

**Conformity Assessment Route:** Annex IX (excluding Chapter II), Regulation EU 2017/745

**Basic UDI-DI:** 619498EC1856X

**Common Specification:** N/A

---

We herewith declare that the above mentioned products are in conformity with the Council Regulation 2017/745 for medical devices.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer. This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

**EC Certificate Number:** G15 049861 0261 Rev. 01

**SRN:** AU-MF-000011753

Signed at Sydney, Australia on: 18 February 2026

DocuSigned by:  
*Nicole Wilson*  
F9F55744DEA64A0...

---

Nicole Wilson  
Person Responsible for Regulatory Compliance (PRRC)  
ResMed Pty. Ltd.

**EC185a.1**  
First issued: 10 December 2021