

ResMed

Lumis™ series

VPAP S

VPAP ST



Home button for returning to the Home screen at any time

Start/Stop button for starting therapy

Dial for navigating and changing options

Air outlet for connecting the air tubing

Quick start guide
English


Read your entire Lumis™ User Guide before use.

Setup



With the device on a stable level surface, grip the retention clip on the back of the device and pull up to open.

Note: The retention clip is shown in the open position.




(a) Plug the power connector into the device power inlet then (b) push down the retention clip to secure in place.

Connect one end of the power cord into the power supply unit and the other end into the power outlet.



Connect the air tubing firmly to the air outlet located on the rear of the device.



Connect the free end of the air tubing firmly onto the assembled mask. Press Start/Stop to begin therapy.

Lumis with HumidAir

If you have a Lumis device with HumidAir™, fill the humidifier as follows:



Open the humidifier and fill with water up to the maximum water level mark.



Close the humidifier and insert it into the side of the device.

Heated humidification offers relief from dryness and congestion and is designed to make therapy more comfortable.

Adjusting to therapy


If you are a first time user, you might need some time to get used to therapy. This is not unusual as it takes most patients between one and two weeks to adjust to the air pressure.

Overcoming symptoms by changing comfort settings

If you are having difficulties getting used to therapy, you can use the Lumis comfort features.

- **Trouble falling asleep**—If you are having trouble falling asleep with high pressure, turn on Ramp or increase Ramp Time.
- **Bloated feeling**—If you are experiencing a slightly bloated feeling from swallowing air, turn on Ramp or increase Ramp Time.
- **Feeling of not getting enough air**—If you feel like you are not getting enough air, turn Ramp Time to Off.


To change the Ramp Time:



1. Press the dial to enter My Options.



2. Turn the dial to highlight Ramp Time and then press to select it.




3. Turn the dial to adjust to your preferred setting.



4. Press the dial to save the change.

If you have a Lumis device with HumidAir, you have more comfort features available which help make your therapy more comfortable.

- **Dry or runny nose**—If you are getting a dry or runny nose, adjust the Humidity Level by turning it up.
- **Droplets of water (condensation)**—If you are getting droplets of water on your nose, mask or air tubing, adjust the Humidity Level by turning it down.



ResMed

Cleaning

1. Wash the air tubing in warm water using mild detergent.
Do not wash in a dishwasher or washing machine.
2. Rinse the air tubing thoroughly and allow to dry out of direct sunlight and/or heat.
3. Wipe the exterior of the device with a dry cloth.

If you have an Lumis device with HumidAir, clean the humidifier as per steps 1 and 2.

Refer to your mask User Guide for detailed instructions on cleaning your mask.

Checking and replacing your parts regularly

It is important for your comfort and health that you check and replace your parts and supplies regularly. Replacing your parts and supplies on a regular basis helps ensure you are receiving optimal therapy and continued comfort.

Refer to your Lumis User Guide for detailed instructions on checking your device. Contact your care provider for information on your replacement schedule.

Serial number	Device number
Mask type and size	Date of setup
Notes	



ResMed Ltd

1 Elizabeth Macarthur Drive
Bella Vista NSW 2153 Australia

DISTRIBUTED BY

ResMed Corp 9001 Spectrum Center Boulevard San Diego CA 92123 USA
EC REP ResMed (UK) Ltd 96 Jubilee Ave Milton Park Abingdon Oxfordshire OX14 4RW UK

See www.resmed.com for other ResMed locations worldwide. For patent information, see www.resmed.com/ip.

Lumis, ClimateLine and HumidAir are trademarks of ResMed Ltd. ClimateLine is registered in U.S. Patent and Trademark Office.

© 2015 ResMed Ltd. 288300/1 2015-03

ResMed.com



0123



288300