



# Non-Vented Mask/Device compatibility list

English

This document specifies the masks compatible with ResMed’s ventilators. For full details regarding the correct use of these devices including settings, please refer to the specific user guide.

Non-vented masks must be used with an active-exhaust-valve ventilator. The ventilators in this list all include an active exhaust valve.

Note: Not all products are available in all regions.

Device	AcuCare F1-0 / F1-1	Quattro Air NV
<b>Astral™</b>		
Astral 100	✓	✓
Astral 150	✓	✓
<b>Elisée</b>		
Elisée 150	✓	✓
Elisée 250	✓	✓
Elisée 350	✓	✓
<b>VS Ultra™</b>		
VS Ultra	✓	✓
VS Integra	✓	✓
<b>VS III™</b>		
VS III	N/V	✓

N/V = The mask has not been validated for this device.



VS III / Elisée ResMed Paris 240 rue de la Motte 77550 Moissy-Cramayel France

Astral ResMed Ltd 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia

**DISTRIBUTED BY**

ResMed Ltd 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia

ResMed Corp 9001 Spectrum Center Boulevard San Diego CA 92123 USA

ResMed (UK) (for Ltd) 96 Jubilee Ave Milton Park Abingdon Oxfordshire OX14 4RW UK

See [www.resmed.com](http://www.resmed.com) for other ResMed locations worldwide. For patent information, see [www.resmed.com/ip](http://www.resmed.com/ip). AcuCare, Quattro and Quattro Air are trademarks of ResMed Ltd. Astral, Elisée, VS Ultra, VS Integra and VS III are trademarks of ResMed Paris SAS. Quattro, Quattro Air and Astral are registered in U.S. Patent and Trademark Office. © 2015 ResMed Ltd. 1017306/5 2015-04

